

# LEGEND MAX™ ELISA Kit



# Mouse IL-34

Cat. No. 439107

ELISA Kit for Accurate Quantitation of Mouse IL-34 from Cell Culture Supernatant, Serum, Plasma and Other Biological Fluids

BioLegend, Inc. biolegend.com



Table of Contents	Page
Introduction	. 2
Materials Provided	3
Materials to be Provided by the End-User	3
Storage Information	. 4
Health Hazard Warnings	. 4
Specimen Collection and Handling	. 5
Reagent and Sample Preparation	. 5
Assay Procedure	6
Assay Procedure Summary	. 8
Calculation of Results	. 9
Typical Data	. 9
Performance Characteristics.	. 9
Troubleshooting Guide	. 11
ELISA Plate Template	. 13

#### Introduction:

IL-34 is a homodimeric glycoprotein which contains 241 amino acids. IL-34 is expressed in heart, brain, lung, liver, kidney, thymus, spleen and other organs. It specifically binds to CD115/M-CSF receptor (CSF-1 receptor or c-Fms) and activates ERK1/2 phosphorylation. The function of IL-34 overlaps with CSF-1 to regulate myeloid lineage differentiation, proliferation, and survival.

Elevated levels of IL-34 are seen in the plasma of rheumatoid arthritis patients, but can be decreased in response to therapy. IL-34 levels are also increased in the serum of cancer patients and associated with poor prognosis.

Mouse IL-34 shares 71% sequence identity with human IL-34 on the amino acid level.

LEGEND MAX™ Mouse IL-34 ELISA kit is a Sandwich Enzyme-Linked Immunosorbent Assay (ELISA) with a 96-well strip plate that is pre-coated with a monoclonal rat anti-mouse IL-34 antibody. The detection antibody is a biotinylated polyclonal goat anti-mouse IL-34. This kit is specifically designed for the accurate quantitation of mouse IL-34 from cell culture supernatant, serum, plasma and other biological fluids. This kit is analytically validated with readyto-use reagents.

#### **Materials Provided:**

Description	Quantity	Volume (per bottle)	Part #
Anti-Mouse IL-34 Pre-coated 96-well Strip Microplate	1 plate		79778
Mouse IL-34 Detection Antibody	1 bottle	12 mL	79779
Mouse IL-34 Standard	1 vial	lyophilized	79781
Avidin-HRP D	1 bottle	12 mL	78237
Assay Buffer A	1 bottle	25 mL	78232
Wash Buffer (20X)	1 bottle	50 mL	78233
Substrate Solution F	1 bottle	12 mL	79132
Stop Solution	1 bottle	12 mL	79133
Plate Sealers	1 pack		78101

# Materials to be Provided by the End-User:

- Microplate reader able to measure absorbance at 450 nm
- Adjustable pipettes to measure volumes ranging from 1  $\mu L$  to 1,000  $\mu L$
- Deionized water
- Wash bottle or automated microplate washer
- Log-Log graph paper or software for data analysis
- Tubes to prepare standard dilutions
- Timer
- Plate Shaker
- Polypropylene vials

# **Storage Information:**

Store unopened kit components between 2°C and 8°C. Do not use this kit beyond its expiration date.

Opened or Reconstituted Components					
Microplate wells	If not all microplate strips are used, remove the excess strips by pressing up from underneath each strip. Place excess strips back in the foil pouch with the included desiccant pack and reseal. Store between 2°C and 8°C for up to one month.				
Standard	The remaining reconstituted standard stock solution ca be aliquoted into polypropylene vials and stored at -70° for up to one month. Avoid repeated freeze-thaw cycles.				
Detection Antibody					
Avidin-HRP D					
Assay Buffer A	Store opened reagents between 2°C and 8°C and use				
Wash Buffer (20X)	within one month.				
Substrate Solution F	]				
Stop Solution					

# **Health Hazard Warnings:**

- 1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin. Refer to the MSDS online at BioLegend's website for details (www.biolegend.com/msds).
- Substrate Solution F is harmful if inhaled or ingested. Avoid skin, eye and clothing contact.
- To reduce the likelihood of blood-borne transmission of infectious agents, handle all serum, plasma and other biological fluids in accordance with NCCLS regulations.
- 4. Stop Solution contains strong acid. *Wear eye, hand, and face protection.*
- 5. Before disposing of the plate, rinse it with an excess amount of tap water.

#### **Specimen Collection and Handling:**

Specimens should be clear and non-hemolyzed. If possible, unknown samples should be run at a number of dilutions to determine the optimal dilution factor that will ensure accurate quantitation. Mouse IL-34 degrades rapidly in biological sample. Fresh samples are recommended.

<u>Cell Culture Supernatant</u>: If necessary, centrifuge all samples to remove debris prior to analysis. It is recommended that samples be stored at < -70°C. Avoid repeated freeze-thaw cycles.

<u>Serum:</u> Use a serum separator tube and allow clotting for at least 30 minutes, then centrifuge for 10 minutes at 1,000 x g. Remove serum layer. Assay immediately or store serum samples at < -70°C. Avoid repeated freeze-thaw cycles.

<u>Plasma:</u> Collect blood samples in citrate, heparin or EDTA containing tubes. Centrifuge for 10 minutes at 1,000 x g within 30 minutes of collection. Assay immediately or store plasma samples at < -70°C. Avoid repeated freeze-thaw cycles.

## **Reagent and Sample Preparation:**


Note: All reagents should be diluted immediately prior to use.

- Dilute the 20X Wash Buffer to 1X with deionized water. For example, make 1 liter of 1X Wash Buffer by adding 50 mL of 20X Wash Buffer to 950 mL of deionized water. If crystals have formed in the 20X Wash Buffer, bring to room temperature and mix until dissolved.
- Reconstitute the lyophilized Mouse IL-34 Standard by adding the volume of Assay Buffer A indicated on the vial label to make the 20 ng/mL standard stock solution. Allow the reconstituted standard to sit at room temperature for 15 minutes, then briefly vortex to mix completely.
- 3. In general, tissue culture supernatant samples are analyzed without dilutions. However, if dilutions are required, samples should be diluted with Assay Buffer A.
- 4. It is recommended that serum and plasma samples be diluted 1:10 with Assay Buffer A prior to analysis.

### **Assay Procedure:**

Note: Do not mix reagents from different kits or lots. Reagents and/or antibodies from different manufacturers should not be used with this kit.

- 1. Bring all reagents to room temperature prior to use. It is strongly recommended that all standards and samples be run in duplicate. A standard curve is required for each assay.
- If not all microplate strips will be used, remove the excess strips by pressing up from underneath each strip. Place excess strips back in the foil pouch with the included desiccant pack and reseal.
- 3. Prepare 500 μL of the 2,000 pg/mL top standard by diluting 50 μL of the standard stock solution in 450 μL of Assay Buffer A. Perform six two-fold serial dilutions of the 2,000 pg/mL top standard in separate tubes using Assay Buffer A as the diluent. Thus, the mouse IL-34 standard concentrations in the tubes are 2,000 pg/mL, 1,000 pg/mL, 500pg/mL, 250 pg/mL, 125 pg/mL, 62.5 pg/mL and 31.3 pg/mL, respectively. Assay Buffer A serves as the zero standard (0 pg/mL).




- 4. Wash the plate 4 times with at least 300  $\mu$ L of 1X Wash Buffer per well and blot any residual buffer by firmly tapping plate upside down on absorbent paper. All subsequent washes should be performed similarly.
- 5. For measuring samples of serum, plasma or cell culture supernatant:
  - a) Add 50 µL of Assay Buffer A to each well that will contain either standard dilutions or samples.
  - b) Add 50  $\mu L$  of standard dilutions or samples to the appropriate wells.
- 6. Seal the plate with a Plate Sealer included in the kit and incubate the plate at room temperature for 2 hours while shaking at 200 rpm.

6

- 7. Discard the contents of the plate into a sink, then wash the plate 4 times with 1X Wash Buffer as in step 4.
- 8. Add 100  $\mu$ L of Mouse IL-34 Detection Antibody solution to each well, seal the plate and incubate at room temperature for 1 hour while shaking.
- 9. Discard the contents of the plate into a sink, then wash the plate 4 times with 1X Wash Buffer as in step 4.
- 10. Add 100  $\mu$ L of Avidin-HRP D solution to each well, seal the plate and incubate at room temperature for 30 minutes while shaking.
- 11. Discard the contents of the plate into a sink, then wash the plate 5 times with 1X Wash Buffer as in step 4. For this final wash, soak wells in 1X Wash Buffer for 30 seconds to 1 minute for each wash. This will help minimize background.
- 12. Add 100  $\mu$ L of Substrate Solution F to each well and incubate for 20 minutes in the dark. Wells containing mouse IL-34 should turn blue in color with an intensity proportional to its concentration. It is not necessary to seal the plate during this step.
- 13. Stop the reaction by adding 100  $\mu$ L of Stop Solution to each well. The solution color should change from blue to yellow.
- 14. Read absorbance at 450 nm immediately. If the reader is capable of reading at 570 nm, the absorbance at 570 nm can be subtracted from the absorbance at 450 nm.

# **Assay Procedure Summary**




#### Calculation of Results:

The data can be best calculated with computer-based curve-fitting software using a 5- or 4-parameter logistics curve-fitting algorithm. If an appropriate software is not available, use log-log graph paper to determine sample concentrations. Determine the mean absorbance for each set of duplicate or triplicate standards, controls, and samples. Plot the standard curve on log-log graph paper with cytokine concentration on the X-axis and absorbance on the Y-axis. Draw a best fit line through the standard points. To determine the unknown cytokine concentrations, find the mean absorbance value of the unknown concentration on the Y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the X-axis and read the cytokine concentration.

If samples were diluted, multiply the concentration by the appropriate dilution factor. If a test sample's absorbance value falls outside the linear portion of the standard curve, the test sample needs to be re-analyzed at a higher (or lower) dilution as appropriate.

### **Typical Data:**

This standard curve was generated at BioLegend for demonstration purposes only. A standard curve must be run with each assay.



#### Performance Characteristics:

<u>Specificity:</u> No cross-reactivity was observed when this kit was used to analyze the following recombinant cytokines/chemokines, each at up to 50 ng/mL.

Mouse	M-CSF, FGF-Basic, GM-CSF, Isthmin1, IL-1α, IL-1β, IL-3, IL-4, IL-6, IL-7, IL-10, IL-12 p40, IL-12 p70, IL-17A, IL-17E/IL-25, IL-21, IL-22, IL-27, IL-33, CXCL10/ IP-10, CXCL13/BLC, CCL2/MCP-1, TNF-α.
Human	IL-34, IL-2, IL-4, IL-12p70, IL-17A, CCL2/MCP-1, IFN-γ, TNF-α.

Sensitivity: The minimum detectable concentration is  $8.4 \pm 1.8 \text{ pg/mL}$  (n=5).

<u>Recovery:</u> Recombinant mouse IL-34, at concentrations of 1000, 250, and 62.5 pg/mL, was spiked into 1:10 dilution of four pooled serum samples, each from a different mouse strain, and then analyzed with the LEGEND MAX<sup>™</sup> Mouse IL-34 ELISA Kit. On average, 95% of the protein was recovered from the serum samples.

<u>Linearity:</u> Four mouse serum samples, each from a different strain, containing high concentration of mouse IL-34, were diluted with Assay Buffer A to levels within the dynamic range, then assayed with the kit to determine the dilution linearity. The linearity of dilution ranged from 94% to 106%. On average, 101% linearity of dilution was observed.

<u>Intra-Assay Precision:</u> Two samples with different mouse IL-34 concentrations were tested with 16 replicates in one assay.

Concentration	Sample 1	Sample 2
Number of Replicates	16	16
Mean Concentration (pg/mL)	563.2	147.0
Standard Deviation	14.4	6.5
% CV	2.6	4.4

<u>Inter-Assay Precision:</u> Two samples with different concentrations of mouse IL-34 were assayed in three independent assays.

Concentration	Sample 1	Sample 2
Number of Assays	3	3
Mean Concentration (pg/mL)	514.4	128.2
Standard Deviation	31.7	12.7
% CV	6.2	9.9

#### Biological Samples:

*Serum*- Five pooled mouse serum samples, each from a different strain, were assayed for basal levels of mouse IL-34. The average detectable concentration was 3450 pg/mL.

Cell Culture Supernatant - Thioglycollate-elicited peritoneal cells from BALB/c mice were stimulated with IFN $\gamma$  and LPS for 22 hours; cell culture supernatant was assayed and contained 16 pg/mL of mouse IL-34. EL-4 cell culture supernatant measured 21 pg/mL of mouse IL-34.

<sup>10</sup> Tel: 877-246-5343

# **Troubleshooting Guide:**

Problem	Probable Cause	Solution		
High Background	Background wells were contaminated	Avoid cross-well contamination by using the provided plate sealers.  Use multichannel pipettes and change tips between pipetting samples and reagents.		
	Insufficient washes	Increase number of washes. Increase soaking time between washes prior to addition of substrate solution.		
	TMB Substrate Solution was contaminated	TMB Substrate Solution should be clear and colorless prior to addition to wells. Use a clean container prior to pipetting substrate solution into wells.		
No or poor signal	Detection Antibody, Avidin-HRP or Substrate solution were NOT added			
	Wrong reagent or reagents were added in wrong sequential order	Rerun the assay and follow the protocol.		
	Insufficient plate agitation	The plate should be agitated during all incubation steps using a plate shaker at a speed where solutions in wells are within constant motion without splashing.		
	The wash buffer contains Sodium Azide (NaN3)	Avoid Sodium Azide contamination in the wash buffer as it inhibits HRP activity.		
	Incubations were done at an inappropriate temperature, timing or without agitation	Rerun the assay and follow the protocol.		
Low or poor standard curve	The standard was incorrectly reconstituted or diluted	Adjust the calculations and follow the protocol.		
signal	Standard was inappropriately stored	Store the reconstituted standard stock solution in polypropylene vials at -70°C. Avoid repeated freeze-thaw cycles.		
	Reagents added to wells with incorrect concentrations	Check for pipetting errors and the correct reagent volume.		

Problem	Probable Cause	Solution		
Signal is high, standard curves have saturated	Standard reconstituted with less volume than required	Reconstitute new lyophilized standard with the correct volume of solution recommended in the protocol.		
signal	Standards/samples, detection antibody, Avidin-HRP or substrate solution were incubated for too long	Rerun the assay and follow the protocol.		
Sample readings	Samples contain no or below detectable levels of the analyte	If samples are below detectable levels, it may be possible to use a larger sample volume. Contact technical support for appropriate protocol modifications.		
are out of range	Samples contain analyte concentrations greater than highest standard point	Samples may require dilution and analysis		
	Multichannel pipette errors	Confirm that pipette calibrations are accurate.		
High variation in	Plate washing was not	Ensure pipette tips are tightly secured.		
samples and/or	adequate or uniform	Ensure uniformity in all wash steps.		
standards	Non-homogenous samples	Thoroughly mix samples before assaying.		
	Samples may have high particulate matter	Remove particulate matter by centrifugation.		
	Cross-well contamination	Do not reuse plate sealers.		
		Always change tips for reagent additions. Ensure that pipette tips do not touch the reagents on the plate.		

	12								
	11								
	10								
	6								
d)	8								
ELISA Plate Template	L								
late Te	9								
LISA P	5								
	4								
	8								
	2								
	1								
		A	В	C	D	E	F	Ð	Н



LEGEND MAX™ Kits are manufactured by **BioLegend Inc.** 

8999 BioLegend Way San Diego, CA 92121 Tel: 1.858.768.5800

Tel US & Canada Toll-Free: 1.877.Bio-Legend (1.877.246.5343)

Fax: 1.877.455.9587

Email: info@biolegend.com

biolegend.com